



TENDER FOR SUPPLY & INSTALLATION OF ANECHOIC CHAMBER
FOR PERMANENT CAMPUS OF THE INSTITUTE AT OKHLA PHASE-
III, NEW DELHI-110020

(Tender no. IIITD/ECE/ Anechoic Chamber/020/2022-23)

AMENDMENT-01

S. No.	Technical Specification	Parameter	Amendment
A.2	Anechoic Chamber Specifications: Anechoic Chamber: 2 nd point	Anechoic Chamber Size (Outer): 3x3x3 Feet (LxWxH) Approximately. Should be upgradable to 110 GHz for considering future aspects. A document in support of this must be enclosed.	Anechoic Chamber Size (Outer): 3m x 3m x 3m (LxWxH) Approximately. Should be upgradable to 110 GHz for considering future aspects. A document in support of this must be enclosed.