

14th September, 2021

Corrigendum

Tender No. IIIT-D/CB/Fluorescence Microscope/012/2021-22

S.no.	As per tender scope of work	New Requirement /Changes
10	Microscope must be compatible with achromatic lens	Microscope must be compatible with achromatic lens and plan fluor; 4X, 10X-achromat and 20X, 40X and 100x should have plan fluor
11	Microscope must support fluorescence LED illumination system with DAPI, FITC, and TRITC filters.	Microscope must support fluorescence LED illumination system with DAPI, FITC, and TRITC filters with more than 20,000 hours or more working life.
17		COMPUTER: BRANDED COMPUTER WITH i7 PROCESSOR, 8 GB RAM, 1 TB HDD